

# Certificate of Analysis



**YSI Incorporated**

*Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA*

*Phone (937) 767-7241 (800) 765-4974 Fax (937) 767-9353*

**Product Model Number:**

3167

**Conductivity Calibration  
Solution**

**Lot Number:**

10A100024

**Testing of Production Standard with a Metrology Calibrated Cell**

**Nominal Conductivity  
(microsiemens/cm):**

1K

**Specification Range,  
(microsiemens/cm):**

995 - 1005

**Test Result,  
(microsiemens/cm):**

998

**Lot Result:**

Accepted

**Inspected by::**

J. Jenkins

**Date Tested:**

14-Jan-10



A handwritten signature in blue ink, appearing to read 'Lisa M. Abel'.

**Lisa M. Abel  
Director, Quality and Corporate  
Responsibility**