Certificate of Analysis

YSI Incorporated
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA
Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category	Membrane
Product Model Number	2329
Lot Number	16L100646
Method	Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	10.0-30.0 nA	13.2 nA
	Average Linearity	14.3-15.7mmol/L	15.14 mmol/L
	Potassium Ferrocyanide	-0.3 - 0.3mmol/L	0.02 mmol/L

Conforms to acceptance criteria

Completed by: Christy Woodford 12/1/2016
Your Name

Lisa M. Abel
Director, Quality and Corporate Responsibility