

# Certificate of Analysis



**YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category            Membrane  
Product Model Number    2365  
Lot Number                 16M100198  
Method                      Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	8.0-45.0 nA	14.0 nA
	Average Linearity	8.55-9.45 g/L	9.00 g/L
	Potassium Ferrocyanide	0.000-0.005 g/L	0.001 g/L

Conforms to acceptance criteria

Completed by:            Nic Heatherington

12/15/2016



A handwritten signature in blue ink, appearing to read 'Lisa M. Abel'.

Lisa M. Abel  
Director, Quality and Corporate  
Responsibility