

# Certificate of Analysis



**YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Bioanalytical Standard  
Product Model Number 2355  
Lot Number 16M100570  
Method Comparative enzymatic testing of Production standard vs Metrology standard

| Chemistry Tested | Determination   | Tolerance           | Result |
|------------------|---|---------------------|--------|
| GLUCOSE          | (Average of Production)-(Average of Metrology)          | $\pm 10\text{mg/L}$ | 7 mg/L |
|                  | Coefficient of variation for production standard (CV,%) | $\leq 2.0\%$        | 0.8 %  |

Conforms to acceptance criteria

Completed by: Kristin Stover

Date: 1/4/2017



A handwritten signature in blue ink, appearing to read 'Lisa M. Abel'.

Lisa M. Abel  
Director, Quality and Corporate  
Responsibility