

# Certificate of Analysis



**YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Membrane  
Product Model Number 2329  
Lot Number 17B100387  
Method Amperometric testing with standards

| Time of Testing | Determination                   | Specification    | Results      |
|-----------------|---------------------------------|------------------|--------------|
| Initial         | Average Calibration Sensitivity | 10.0-30.0 nA     | 19.8 nA      |
|                 | Average Linearity               | 14.3-15.7mmol/L  | 14.98 mmol/L |
|                 | Potassium Ferrocyanide          | -0.3 - 0.3mmol/L | 0.01 mmol/L  |

Conforms to acceptance criteria

Completed by: Christy Woodford  
Your Name

3/15/2017



Lisa M. Abel  
Director, Quality and Corporate  
Responsibility