

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Bioanalytical Standard
Product Model Number 2355
Lot Number 17B100545
Method Comparative enzymatic testing of Production standard vs Metrology standard

Chemistry Tested	Determination	Tolerance	Result
GLUCOSE	(Average of Production)-(Average of Metrology)	$\pm 10\text{mg/L}$	7 mg/L
	Coefficient of variation for production standard (CV,%)	$\leq 2.0\%$	0.9 %

Conforms to acceptance criteria

Completed by: Kristin Stover

Date: 3/2/2017



A handwritten signature in blue ink, appearing to read 'Lisa M. Abel'.

Lisa M. Abel
Director, Quality and Corporate
Responsibility