

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category	Membrane
Product Model Number	2329
Lot Number	17D100515
Method	Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	10.0-30.0 nA	12.6 nA
	Average Linearity	14.3-15.7mmol/L	14.69 mmol/L
	Potassium Ferrocyanide	-0.3 - 0.3mmol/L	0.00 mmol/L

Conforms to acceptance criteria

Completed by: Christy Woodford 4/25/2017
 Your Name



Lisa M. Abel
 Director, Quality and Corporate
 Responsibility