Certificate of Analysis

YSI Incorporated
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA
Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category: Buffer and solid products

Product Model Number: 2357

Lot Number: 190925GFS098005

Method: Verification of mass and pH (for buffers)

Determination	Specification	Result
Average mass	6.00 – 6.70 g	6.17g
Average pH (buffers only)	7.1-7.5	7.3

Conforms to acceptance criteria

Completed by: David Stoler Date: 10/16/2019

Lisa M. Abel
Director, Quality and Corporate Responsibility