

# Certificate of Analysis



YSI Incorporated  
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA  
Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category                      Bioanalytical Standard  
Product Model Number                      2772  
Lot Number                                      19J100139  
Method    Comparative enzymatic testing of Production standard vs. Metrology standard

<b>Chemistry Tested</b>	<b>Determination</b>	<b>Tolerance</b>	<b>Result</b>
CHOLINE	(Average of production) - (average of metrology)	$\pm 2$ mg/L	1.35 mg/L
	Coefficient of variation for production standard (CV, %)	$\leq 2\%$	0.6%
	(Average of production) - (average of metrology)	N/A	N/A
	Coefficient of variation for production standard (CV, %)	N/A	N/A

Conforms to acceptance criteria

Completed by: Lucas Stover    Date: 09/12/2019



A handwritten signature in blue ink, appearing to read "Lisa M. Abel".

Lisa M. Abel  
Director, Quality and Corporate  
Responsibility