

# Certificate of Analysis



**YSI Incorporated**  
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA  
Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category                    Bioanalytical Standard  
Product Model Number                7141  
Lot Number                                19J100586  
Method                                    Comparative enzymatic testing of Production standard vs. Metrology standard

Chemistry Tested	Determination	Tolerance	Result
Glycerol	(Average of production) - (average of metrology)	$\pm 0.30$ g/L	0.07g/L
	Coefficient of variation for production standard (CV, %)	$\leq 2$ %	0.98 %
None	(Average of production) - (average of metrology)	N/A	N/A
	Coefficient of variation for production standard (CV, %)	N/A	N/A

Conforms to acceptance criteria

Completed by: Lucas Stover

Date: 09/20/2019



Lisa M. Abel  
Director, Quality and Corporate  
Responsibility