Certificate of Analysis

YSI Incorporated
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA
Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category ISE Standard

Product Model Number 7179

Lot Number 19K100450

Method Comparative ISE testing of Production standard vs. Metrology standard

Chemistry Tested	Determination	Tolerance	Result
AMMONIUM	(Average of production) - (average of metrology)	<0.250 mV	0.001 mV
	Coefficient of variation for production standard (CV, %)	≤2%	0.1%
POTASSIUM	(Average of production) - (average of metrology)	<0.250 mV	0.084 mV
	Coefficient of variation for production standard (CV, %)	≤2%	0.1%

Conforms to acceptance criteria

Completed by: Lucas Stover Date: 10/23/2019

Lisa M. Abel
Director, Quality and Corporate Responsibility