

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Membrane
Product Model Number 2735
Lot Number 19L100366
Method Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	8.0-40.0 nA	10.2 nA
	Average Linearity	7.6-8.4mmol/L	7.87 mmol/L
	Potassium Ferrocyanide	< 0.4mmol/L	0.05 mmol/L

Conforms to acceptance criteria

Completed by:

Christy Woodford

11/26/2019

Your Name



A handwritten signature in blue ink that reads 'Lisa M. Abel'.

Lisa M. Abel
Director of Quality