Certificate of Analysis

YSI Incorporated
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA
Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

Product Category: Bioanalytical Standard

Product Model Number: 2801

Lot Number: 19M100074

Method: Comparative enzymatic testing of Production standard vs. Metrology standard

Chemistry Tested	Determination	Tolerance	Result
GLUCOSE	(Average of production) - (average of metrology)	± 5.0 mg/L	1.15 mg/L
	Coefficient of variation for production standard (CV, %)	≤ 2.0%	0.97 %

Conforms to acceptance criteria

Completed by: David Stoler Date: 12-5-2019

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.

ISO 9001 CERTIFIED YSI INCORPORATED

Gregory W. Popp
Quality Manager, Analytics