

# Certificate of Analysis



**YSI Incorporated**  
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA  
Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

a xylem brand

---

Product Category                      Bioanalytical Standard  
Product Model Number                2778  
Lot Number                                20B130085  
Method                                      Comparative enzymatic testing of Production standard vs. Metrology standard

Chemistry Tested	Determination	Tolerance	Result
Sucrose	(Average of production) - (average of metrology)	$\pm 0.3$ g/L	0.05 g/L
	Coefficient of variation for production standard (CV, %)	$\leq 2.0\%$	0.56%

---

Conforms to acceptance criteria

Completed by: David Stoler

Date: 3-6-2020

*The procedures used to manufacture and test this product have been verified by QA for GMP compliance.*



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp  
Quality Manager, Analytics