

Certificate of Analysis



YSI Incorporated
 Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA
 Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

a xylem brand

Product Category	Buffer and solid products
Product Model Number	1579
Lot Number	20L100039
Method	Verification of mass and pH (for buffers)

Determination	Specification	Result
Average mass	N/A	
Average pH (buffers only)	9.00 ± 0.10	9.00

Conforms to acceptance criteria

Completed by: Nick Rittenhouse Date: 11/4/2020
Amanda Fenner

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Gregory W. Popp
 Quality Manager, Analytics