

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Buffer and solid products
Product Model Number 2363
Lot Number 21J100278
Method Verification of mass

Determination	Specification	Result
Average Mass	1.15-1.35 grams	1.25 g

Conforms to acceptance criteria

Completed by: Amanda Fenner

Date: 9/14/2021



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp

Quality Manager Analytics