

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Bioanalytical Standard
Product Model Number 1530
Lot Number 22B100319
Method Comparative enzymatic testing of Production standard vs Metrology standard

Chemistry Tested	Determination	Tolerance	Result
LACTATE	(Average of Production)-(Average of Metrology)	$\pm 0.05\text{g/L}$	0.00 g/L
	Coefficient of variation for production standard (CV,%)	$\leq 2.0\%$	0.9 %

Conforms to acceptance criteria

Completed by: Amanda Fenner

Date: 2/16/2022

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp
Quality Manager, Analytics