

# Certificate of Analysis



**YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category      Membrane  
Product Model Number    2365  
Lot Number              22D100002  
Method                  Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	8.0-45.0 nA	11.5 nA
	Average Linearity	8.55-9.45 g/L	8.95 g/L
	Potassium Ferrocyanide	0.00-0.05	g/L

Conforms to acceptance criteria

Completed by:      Sandy Crawford  
                            Your Name

4/1/2022



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp  
Quality Manager Analytics