



# Certificate of Analysis

## Bioanalytical Standard

### 2747

1725 Brannum Lane  
Yellow Springs, Ohio  
45387-1107 USA  
+1 937-767-7241

Method:

Comparative enzymatic testing of Production standard vs Metrology standard

## Testing Results

Lot Number: 22J100181

Chemistry Tested	Determination	Tolerance	Result
GLUCOSE	(Average of Production)- (Average of Metrology)	$\pm 0.01$ g/L	0.002 g/L
	Coefficient of variation for production standard (CV,%)	$\leq 2.0\%$	0.62 %
LACTATE	(Average of Production)- (Average of Metrology)	$\pm 0.01$ g/L	0.001 g/L
	Coefficient of variation for production standard (CV,%)	$\leq 2.0\%$	0.44 %

---

Conforms to acceptance criteria

Completed by: afenner Date: 9/16/2022

*The procedures used to manufacture and test this product have been verified by QA for GMP compliance.*