



Certificate of Analysis

Bioanalytical Standard

2356

1725 Brannum Lane
Yellow Springs, Ohio
45387-1107 USA
+1 937-767-7241

Method:

Comparative enzymatic testing of Production standard vs Metrology standard

Testing Results

Lot Number: 22J100384

| Chemistry Tested | Determination | Tolerance | Result |
|------------------|--|-----------|-----------|
| GLUCOSE | (Average of Production)- (Average of Metrology) | ±25 mg/L | 3.13 mg/L |
| | Coefficient of variation for production standard (CV,%) | ≤2.0% | 0.5 % |

Conforms to acceptance criteria

Completed by: afenner Date: 10/5/2022

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.