



Certificate of Analysis

Bioanalytical Standard

2747

1725 Brannum Lane
Yellow Springs, Ohio
45387-1107 USA
+1 937-767-7241

Method:

Comparative enzymatic testing of Production standard vs Metrology standard

Testing Results

Lot Number: 22K100113

Chemistry Tested	Determination	Tolerance	Result
GLUCOSE	(Average of Production)- (Average of Metrology)	± 0.01 g/L	0.002 g/L
	Coefficient of variation for production standard (CV,%)	$\leq 2.0\%$	0.65 %
LACTATE	(Average of Production)- (Average of Metrology)	± 0.01 g/L	0.002 g/L
	Coefficient of variation for production standard (CV,%)	$\leq 2.0\%$	0.55 %

Conforms to acceptance criteria

Completed by: NRITTEN Date: 10/4/2022

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.