



**Certificate of Analysis**  
**Bioanalytical Standard**  
**2327**

1725 Brannum Lane  
Yellow Springs, Ohio  
45387-1107 USA  
+1 937-767-7241

Method:

Comparative enzymatic testing of Production standard vs Metrology standard

## Testing Results

Lot Number: 22K100517

Chemistry Tested	Determination	Tolerance	Result
LACTATE	(Average of Production)- (Average of Metrology)	±0.1 mmol/L	0.03 mmol/L
	Coefficient of variation for production standard (CV,%)	≤2.0%	0.6 %

---

Conforms to acceptance criteria

Completed by: NRITTEN Date: 10/26/2022

*The procedures used to manufacture and test this product have been verified by QA for GMP compliance.*