Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

Product Category

Buffer and solid products

Product Model Number

2357

Lot Number

230302GFS098005

Method

Verification of mass and pH (for buffers)

Determination	Specification	Result
Average mass	6.00 – 6.70 g	6.17 g
Average pH (buffers only)	7.1-7.5	7.3

Conforms to acceptance criteria

Completed by:

Nick Rittenhouse

Date:

3/29/2023 3/29/2023

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Gregory W. Popp Quality Manager, Analytics

Dregory W. Popp