

Certificate of Analysis





YSI Incorporated
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA
Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

a xylem brand


Product Category Buffer and solid products
Product Model Number 2357
Lot Number 230316GFS098005
Method Verification of mass and pH (for buffers)

Determination	Specification	Result
Average mass	6.00 – 6.70 g	6.18 g
Average pH (buffers only)	7.1-7.5	7.3

Conforms to acceptance criteria

Completed by: Nick Rittenhouse  Date: 3/29/2023 

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Gregory W. Popp
Quality Manager, Analytics