

Certificate of Analysis



YSI Incorporated
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA
Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

a xylem brand

Product Category Buffer and solid products
Product Model Number 2357
Lot Number 230502GFS098005
Method Verification of mass and pH (for buffers)

| Determination | Specification | Result |
|---------------------------|---------------|--------|
| Average mass | 6.00 – 6.70 g | 6.21 g |
| Average pH (buffers only) | 7.1-7.5 | 7.2 |

Conforms to acceptance criteria

Completed by:

Amanda Fenner

Handwritten signature of Amanda Fenner in black ink.

Date:

5/22/2023

5/22/2023

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Handwritten signature of Gregory W. Popp in black ink.

Gregory W. Popp
Quality Manager, Analytics