

# Certificate of Analysis



YSI Incorporated  
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA  
Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

---

Product Category	Buffer and solid products
Product Model Number	<u>2357</u>
Lot Number	<u>230912GFS098005</u>
Method	Verification of mass and pH (for buffers)

---

Determination	Specification	Result
Average mass	6.00 – 6.70 g	6.22 g
Average pH (buffers only)	7.1-7.5	7.3

---

Conforms to acceptance criteria

Completed by: Nick Rittenhouse Date: 9/19/2023

*The procedures used to manufacture and test this product have been verified by QA for GMP compliance.*



A blue ink signature that reads 'Gregory W. Popp'.

Gregory W. Popp  
Quality Manager, Analytics