



# Certificate of Analysis

1725 Brannum Lane  
Yellow Springs, Ohio  
45387-1107 USA  
+1 937-767-7241

Product Category            ISE Reagent  
Product Model Number    2970 (10 mg/L Ammonium, 10 mg/L Potassium)  
Lot Number                23A100473  
Method                      Comparative ISE testing of Production reagent vs. Metrology reagent

## Testing Results

Chemistry Tested	Determination	Tolerance	Result
Ammonium	(Average of Production) – (Average of Metrology)	$\leq 0.2$ mg/L (2%)	0.00
	Coefficient of Variation (CV) for Production reagent	$\leq 5\%$	0.35
Potassium	(Average of Production) – (Average of Metrology)	$\leq 0.3$ mg/L (3%)	0.05
	Coefficient of Variation (CV) for Production reagent	$\leq 5\%$	1.70

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.

Technician

Date

Travis Kidder

1/25/2023