

# Certificate of Analysis



**YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category      Membrane  
Product Model Number    2754  
Lot Number                23A100578  
Method                     Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	8.0-40.0 nA	09.4 nA
	Average Linearity	9.50-10.50mmol/L	9.89 mmol/L
	Potassium Ferrocyanide	< 0.4mmol/L	0.05 mmol/L

Conforms to acceptance criteria

Completed by:

Christy Woodford  
Your Name

1/31/2023



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp  
Quality Manager Analytics