

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Membrane
Product Model Number 2725
Lot Number 23F100044
Method Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	5.0-25.0 nA	6.41 nA
	Average Linearity	2.375-2.625 g/L	2.583 g/L
	Potassium Ferrocyanide	-0.2-0.2 g/L	0.00 g/L
Final	Average Calibration Sensitivity	5.0-25.0 nA	7.87 nA
	Average Linearity	2.375-2.625 g/L	2.503 g/L
	Potassium Ferrocyanide	-0.2-0.2 g/L	0.00 g/L

Conforms to acceptance criteria

Completed by: _____ Christy Woodford
Your Name

6/12/2023



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp
Quality Manager