

# Certificate of Analysis



**YSI Incorporated**  
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA  
Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

Product Category                      Membrane  
Product Model Number                2702  
Lot Number                                23H100235  
Method                                      Amperometric testing with standards

Time of Testing	Determination	Specification	Result
Initial	Average calibration sensitivity	10 –30 nA	12.00 nA
	Average Linearity	23.75 – 26.25 g/L	25.44 g/L
	Potassium Ferrocyanide interference	N/A	N/A
Day = _____	Average calibration sensitivity	N/A	N/A
	Average Linearity	N/A	N/A
	Potassium Ferrocyanide interference	N/A	N/A

Conforms to acceptance criteria

Completed by:                                      Christy Woodford                                      Date:                                      8/15/2023

*The procedures used to manufacture and test this product have been verified by QA for GMP compliance.*



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp  
Quality Manager, Analytics