

Certificate of Analysis



YSI Incorporated
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA
Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

a xylem brand

Product Category Membrane
Product Model Number 7140
Lot Number 23J100172
Method Amperometric testing with standards

| Time of Testing | Determination | Specification | Result |
|-----------------|-------------------------------------|---------------|-----------|
| Initial | Average calibration sensitivity | 10 – 75 nA | 23.97 nA |
| | Average Linearity | 38 – 42 g/L | 39.16 g/L |
| | Potassium Ferrocyanide interference | N/A | N/A |
| Day = _____ | Average calibration sensitivity | N/A | N/A |
| | Average Linearity | N/A | N/A |
| | Potassium Ferrocyanide interference | N/A | N/A |

Conforms to acceptance criteria

Completed by: Christy Woodford Date: 9/12/2023

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp
Quality Manager, Analytics