

# Certificate of Analysis



**YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category            Membrane  
Product Model Number    2365  
Lot Number                 23J100174  
Method                      Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	8.0-45.0 nA	12.8 nA
	Average Linearity	8.55-9.45 g/L	8.90 g/L
	Potassium Ferrocyanide	0.00-0.05	0.00 g/L

Conforms to acceptance criteria

Completed by:            Sandy Crawford  
                                  Your Name

9/8/2023



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp  
Quality Manager Analytics