

# Certificate of Analysis



**YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category            Membrane  
Product Model Number    2754  
Lot Number                    23K100626  
Method                         Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	8.0-40.0 nA	10.4 nA
	Average Linearity	9.50-10.50mmol/L	9.92 mmol/L
	Potassium Ferrocyanide	< 0.4mmol/L	0.02 mmol/L

Conforms to acceptance criteria

Completed by: \_\_\_\_\_ Christy Woodford

10/31/2023

Your Name



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp  
Quality Manager Analytics