Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

Product Category Buffer and solid products

Product Model Number 2357

Lot Number 240715GFS098005

Method Verification of mass and pH (for buffers)

Determination	Specification	Result
Average mass	6.00 - 6.70 g	6.35 g
Average pH (buffers only)	7.1 - 7.5	7.3

The procedures used to manufacture and test this product have been verified by QA for compliance. The electronic signature below is representative of the Chemistry Technician testing and verifying conformity to acceptance criteria.

Completed by: Amanda Fenner Date: 8/01/2024

