

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Membrane
Product Model Number 2754
Lot Number 24B100014
Method Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	8.0-40.0 nA	10.5 nA
	Average Linearity	9.50-10.50mmol/L	9.84 mmol/L
	Potassium Ferrocyanide	< 0.4mmol/L	0.02 mmol/L

The procedures used to manufacture and test this product have been verified by QA for compliance. The electronic signature below is representative of the Chemistry Technician testing and verifying conformity to acceptance criteria.

Completed by: Sandy Crawford

Your Name

A handwritten signature in cursive script that reads "Sandy Crawford".

Date: 2/5/2024

