

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Membrane
Product Model Number 2771
Lot Number 24B100406
Method Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	>5 nA	11.33nA
	Average Linearity	427 - 473 mg/L	449.50mg/L
	Potassium Ferrocyanide Interference	<15 mg/L	0.00mg/L
Day 2	Average Calibration Sensitivity	>5 nA	11.23nA
	Average Linearity	427 - 473 mg/L	434.90mg/L
	Potassium Ferrocyanide Interference	<15 mg/L	0.00mg/L

Conforms to acceptance criteria

Completed by:

Sandy Crawford

2/19/2024

Your Name

Sandy Crawford



Gregory W. Popp

Gregory W. Popp
Quality Manager Analytics