

# Certificate of Analysis



**YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category            Membrane  
Product Model Number      7140  
Lot Number                    24B100518  
Method                         Amperometric testing with standards

Time of Testing	Determination	Specification	Result
Initial	Average calibration sensitivity	10 – 75 nA	27.07 nA
	Average Linearity	38 – 42 g/L	38.8 g/L
	Potassium Ferrocyanide interference	N/A	N/A
Day = <u>N/A</u>	Average calibration sensitivity	N/A	N/A
	Average Linearity	N/A	N/A
	Potassium Ferrocyanide interference	N/A	N/A

Conforms to acceptance criteria

Completed by:    Christy Woodford *Christy Woodford*    Date:    2/27/2024



*Lisa M. Abel*  
Lisa M. Abel  
Director, Quality and Corporate  
Responsibility