Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category	Membrane
Product Model Number	2702
Lot Number	24D100109
Method	Amperometric testing with standards

Time of Testing	Determination	Specification	Result
Initial	Average calibration sensitivity	10-30 nA	17.25 nA
	Average Linearity	23.75 – 26.25 g/L	24.62 g/L
	Potassium Ferrocyanide interference	N/A	N/A
Day = N/A	Average calibration sensitivity	N/A	N/A
	Average Linearity	N/A	N/A
	Potassium Ferrocyanide interference	N/A	N/A

Conforms to acceptance criteria

Completed by:

Christy Woodford

Date:

04/11/2024



mullel

Lisa M. Abel Director, Quality and Corporate Responsibility