## **Certificate of Analysis**



**YSI Incorporated** 

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category	Membrane
------------------	----------

Product Model Number 2702

Lot Number 24D100321

Method Amperometric testing with standards

Time of Testing	Determination	Specification	Result
Initial	Average calibration sensitivity	10 - 30 nA	15.71 nA
	Average Linearity	23.75 – 26.25 g/L	25.12 g/L
	Potassium Ferrocyanide interference	N/A	N/A
Day = N/A	Average calibration sensitivity	N/A	N/A
	Average Linearity	N/A	N/A
	Potassium Ferrocyanide interference	N/A	N/A

Conforms to acceptance criteria						
Completed by:	Christy Woodford	Date:	4/16/2024			



Lisa M. Abel Director, Quality and Corporate Responsibility

Thralel