Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

Product Category	Membrane
Product Model Number	2702
Lot Number	24E100552
Method	Amperometric testing with standards

Time of Testing Specification Determination Result Initial Average calibration sensitivity 10 - 30 nA14.71 nA 23.75-26.25 g/L Average Linearity 25.46 g/L Potassium Ferrocyanide interference N/A N/A Average calibration sensitivity N/A N/A Day = Average Linearity N/A N/A Potassium Ferrocyanide interference N/A N/A

The procedures used to manufacture and test this product have been verified by QA for compliance. The electronic signature below is representative of the Chemistry Technician testing and verifying conformity to acceptance criteria.

Completed by:

Christy Woodford

Date: 5/29/2024

