

# Certificate of Analysis



**YSI Incorporated**  
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA  
Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category            Membrane  
Product Model Number    2771  
Lot Number                 24F100304  
Method                      Amperometric testing with standards

Time of Testing	Determination	Specification	Result
Initial	Average calibration sensitivity	>5 nA	12.85 nA
	Average Linearity	427-473mg/L	440.35
	Potassium Ferrocyanide interference	<15mg/L	0.00 mg/L
Day = <u>2</u>	Average calibration sensitivity	>5nA	12.65
	Average Linearity	427-473 mg/L	438.55
	Potassium Ferrocyanide interference	<15mg/L	0.22

Conforms to acceptance criteria

Completed by: Sandy Crawford

Date: 6/17/2024



Ron A. Johnson  
Director of Quality Systems