



# Certificate of Analysis

1725 Brannum Lane  
Yellow Springs, Ohio  
45387-1107 USA  
+1 937-767-7241

Product Category            ISE Reagent  
Product Model Number    2972 (500 mg/L Ammonium, 10 mg/L Potassium)  
Lot Number                 24G100527  
Method                      Comparative ISE testing of Production reagent vs. Metrology reagent

## Testing Results

Chemistry Tested	Determination	Tolerance	Result
Ammonium	(Average of Production) – (Average of Metrology)	$\leq 10$ mg/L (2%)	2.40
	Coefficient of Variation (CV) for Production reagent	$\leq 3\%$	0.89

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.

Technician

Amanda Fenner

Date

7/23/2024

7/23/2024