

# Certificate of Analysis



**YSI Incorporated**  
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA  
Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category            Membrane  
Product Model Number    2701  
Lot Number                 24G100638  
Method                      Amperometric testing with standards

Time of Testing	Determination	Specification	Result
Initial	Average calibration sensitivity	10-50nA	10.75nA
	Average Linearity	N/A	N/A
	Potassium Ferrocyanide interference	+/-1.0nA	0 mg/L
	Average calibration sensitivity	N/A	N/A
	Average Linearity	N/A	N/A
	Potassium Ferrocyanide interference	N/A	N/A

Conforms to acceptance criteria

Completed by: Sandy Crawford

Date: 7/26/2024



Ron A. Johnson  
Director of Quality Systems