

Certificate of Analysis



YSI Incorporated
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA
Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Membrane
Product Model Number 7140
Lot Number 24K100052
Method Amperometric testing with standards

Time of Testing	Determination	Specification	Result
Initial	Average calibration sensitivity	10 - 75 nA	26.13 nA
	Average Linearity	38 - 42 g/L	38.1 g/L
	Potassium Ferrocyanide interference	N/A	N/A
Day = <u>NA</u>	Average calibration sensitivity	N/A	N/A
	Average Linearity	N/A	N/A
	Potassium Ferrocyanide interference	N/A	N/A

Conforms to acceptance criteria

Completed by:

Sandy Crawford

Date:

10/8/2024



Ron A. Johnson
Director of Quality Systems