

# Certificate of Analysis



**YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

a xylem brand

Product Category            Check Solution  
Product Model Number      7173  
Lot Number                    24K100600  
Method                         Comparative conductivity testing of Production standard vs. Metrology standard

Chemistry Tested	Determination	Tolerance	Result
Sodium	(Average of production) – (average of metrology)	$\pm 1.0 \%$	0.32 %
	Coefficient of variation for production standard (CV, %)	$\leq 2.0 \%$	0.43 %

*The procedures used to manufacture and test this product have been verified by QA for compliance. The electronic signature below is representative of the Chemistry Technician testing and verifying conformity to acceptance criteria.*

Completed by: Amanda Fenner

A handwritten signature in black ink, appearing to read 'Amanda Fenner'.

Date: 11/01/2024

A handwritten date in black ink, '11/01/2024'.

