



Certificate of Analysis

1725 Brannum Lane
Yellow Springs, Ohio
45387-1107 USA
+1 937-767-7241

Product Category ISE Reagent
Product Model Number 2971 (1000 mg/L Potassium, 10 mg/L Ammonium)
Lot Number 24L100558
Method Comparative ISE testing of Production reagent vs. Metrology reagent

Testing Results

Chemistry Tested	Determination	Tolerance	Result
Potassium	(Average of Production) – (Average of Metrology)	≤ 20 mg/L (2%)	4.08
	Coefficient of Variation (CV) for Production reagent	≤ 3%	0.95

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.

Technician

Amanda Fenner

Date

12/5/2024

12/05/2024