

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

a xylem brand

Product Category	Buffer and solid products
Product Model Number	<u>2705</u>
Lot Number	<u>251125GFS002699</u>
Method	Verification of mass and pH (for buffers)

Determination	Specification	Result
Average mass	$6.35 \pm 0.35\text{g}$	6.17g
Average pH (buffers only)	7.3 ± 0.2	7.3

The procedures used to manufacture and test this product have been verified by QA for compliance. The electronic signature below is representative of the Chemistry Technician testing and verifying conformity to acceptance criteria.

Completed by: Nick Rittenhouse Date: 12/3/2025

