

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Membrane
Product Model Number 2754
Lot Number 25B100146
Method Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	8.0-40.0 nA	10.4 nA
	Average Linearity	9.50-10.50mmol/L	10.01 mmol/L
	Potassium Ferrocyanide	< 0.4mmol/L	0.01 mmol/L

The procedures used to manufacture and test this product have been verified by QA for compliance. The electronic signature below is representative of the Chemistry Technician testing and verifying conformity to

Completed by:

Christy Woodford
Christy Woodford

Date: 2/11/2025

Your Name

