

# Certificate of Analysis



**YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

---

Product Category	Buffer and solid products
Product Model Number	2392C
Lot Number	25B100581
Method	Verification of mass

---

Determination	Specification	Result
Average Mass	0.20-0.30 grams	0.25 g

---

The procedures used to manufacture and test this product have been verified by QA for compliance. The electronic signature below is representative of the Chemistry Technician testing and verifying conformity to acceptance criteria.

Completed by: \_\_\_\_\_

Date: 2/27/2025

