

# Certificate of Analysis



**YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category      Membrane  
Product Model Number    2329C  
Lot Number              25C100039  
Method                  Amperometric testing with standards

| Time of Testing | Determination                   | Specification    | Results |
|-----------------|---------------------------------|------------------|---------|
| Initial         | Average Calibration Sensitivity | 6.0-30.0 nA      | nA      |
|                 | Average Linearity               | 14.3-15.7mmol/L  | mmol/L  |
|                 | Potassium Ferrocyanide          | -0.3 - 0.3mmol/L | mmol/L  |

*The procedures used to manufacture and test this product have been verified by QA for compliance. The electronic signature below is representative of the Chemistry Technician testing and verifying conformity to acceptance criteria.*

Completed by: Sandy Crawford

Date: 3/5/2025

